



GLYCEROL

DETERMINATION OF GLYCEROL IN SERUM, PLASMA AND FOODSTUFFS




- Colorimetric Method
- Instrument Application Sheets Available
- Use Serum, Plasma or Foodstuffs
- Incl. Glycerol Standard
- Wavelength 546 nm



Products	Product no.	Quantity
Glycerol Reagent Set	2913	50 - 200 tests

SUMMARY

ASSAY PRINCIPLE



SAMPLE MATERIAL

Serum, heparinized or EDTA-plasma.

Stability of the sample: 48 hours at room temperature,
7 days at 2-6°C.

LINEARITY

Up to 11.40 mmol/l Glycerol.

EXPECTED VALUES

Serum, plasma: 0.03 – 0.19 mmol/l Glycerol

NOTES

1. For in vitro diagnostic use only.
2. For professional use only.
3. Always contact INstruChemie for the complete product insert and latest edition.
4. Printed in the Netherlands, Glycerol-summary-280729-1.FEN