



HbA1c CALIBRATOR

AS CALIBRATOR FOR THE DETERMINATION OF GLYCOSYLATED HAEMOGLOBINS FOR ALL METHODS IN WHOLE BLOOD SAMPLES.

- Independent Calibrator
- Expiration date > 5 years, after production
- Lyophilised



Products	Product no.	Quantity
HbA1c Calibrator Low Level	5098	10 x 200 µl
HbA1c Calibrator Medium Level	5099	10 x 200 µl
HbA1c Calibrator High Level	5100	10 x 200 µl

SUMMARY

COMPOSITION

Lyophilised human blood preparation containing a low, medium or high glycosylated haemoglobins. The amount of freeze dried material is corresponding with 0.2 ml whole blood.

APPLICATION

As low, medium or high calibrator for the determination of glycosylated haemoglobins for all methods in whole patient samples. The assigned values are comparable with the DCCT study or the IFCC reference method.

ANALYTICAL RESULTS

Article nr.: 5098, HbA1c Calibrator, Low Contents: 10 x 200 µl

Component	Method	DCCT	IFCC
HbA1c	All Methods	5.1%	3.2

Article nr.: 5099, HbA1c Calibrator, Medium Contents: 10 x 200 µl

Component	Method	DCCT	IFCC
HbA1c	All Methods	8.1%	6.5

Article nr.: 5100, HbA1c Calibrator, High Contents: 10 x 200 µl

Component	Method	DCCT	IFCC
HbA1c	All Methods	10.08%	8.67

NOTES

1. For in vitro diagnostic use only.
2. For professional use only.
3. Always contact INstru[®]chemie for the complete product insert and latest edition.
4. Printed in the Netherlands, HbA1c-Calibrators -summary-280908-1.FEN