



## HbA1c CONTROL

AS CONTROL FOR ACCURACY AND PRECISION FOR THE DETERMINATION OF GLYCOSYLATED HAEMOGLOBINS FOR ALL METHODS IN WHOLE BLOOD PATIENT SAMPLES.

- Independent Control
- Expiration date > 5 years, after production
- Lyophilised



Products	Product no.	Quantity
HbA1c Control Normal Level	2884	10 x 200 µl
HbA1c Control High Level	2887	10 x 200 µl

## SUMMARY

### COMPOSITION

Lyophilised human blood preparation containing a normal or high glycosylated haemoglobins. The amount of freeze dried material is corresponding with 0.2 ml whole blood.

### APPLICATION

As normal or high quality control for the determination of glycosylated haemoglobins for all methods in whole patient samples. The assigned values are comparable with the DCCT study or the IFCC reference method.

### ANALYTICAL RESULTS

Article nr.: 2884, HbA1c Control, Normal Contents: 10 x 200 µl

Component	Method	DCCT	Range	IFCC	Range
HbA1c	All Methods	5.1%	4.4-5.8	3.2%	2.7-3.7

Article nr.: 2887, HbA1c Control, High Contents: 10 x 200 µl

Component	Method	DCCT	Range	IFCC	Range
HbA1c	All Methods	10.1%	8.5-11.7	8.7%	7.3-10.1

### NOTES

1. For in vitro diagnostic use only.
2. For professional use only.
3. Always contact INstruChemie for the complete product insert and latest edition.
4. Printed in the Netherlands, HbA1c-Controls -summary-280908-1.FEN